

CERTIFICATE OF ANALYSIS

Certificate of Analysis

Product name: PBSS (Phosphate Buffered Serology Saline)

Product code(s): 04-004

Batch Number(s): 01180992

Date Manufactured: 05 Sep 2018

Expiry Date: 04 Sep 2020

Appearance: Clear Liquid

Quality Control:

	Specification	This batch result
pH	6.95-7.05	7.02
Conductivity($\mu\text{s}/\text{cm}$)	N/A	N/A
Osmolality(mOsm/kg)	285-295	286

Meets the requirements set out in 'Guidelines for the Blood Transfusion Services in the United Kingdom' 7th Edition; as set out in chapter 12, 12.2.4, Miscellaneous reagents.

Preservatives used: Sodium azide (0.01% [w/v])

Sterility: Product is filtered through a 0.2 μm filter into gamma irradiated bags

Storage: Room temperature

Country of Origin: United Kingdom

Health and Safety:

Precautionary Notes

Please download the Safety Data Sheet at <http://sourcebioscience.com>

Signed for Quality Control: D Crosste .

Date: 07 - Sept - 18

Signed for Certificate Check: o~Q~u.

Date: 07 SEP 18

Author: Rafael Carlisle
Technical Reviewer: Paul Booth
QA Reviewer: Jahira Jangaria

SER-DOC-32 v1.0
Authorised Date: 06 Jul 17
Next Review Date: 06 Jul 18