

CERTIFICATE OF ANALYSIS

Certificate of Analysis

Product name: Reagent Grade Water

Product code(s): 04-035

Batch Number(s): 18180995

Date Manufactured: 15 Oct 2018

Expiry Date: 14 Oct 2019

Appearance: Clear liquid

Quality Control:

	Specification	This batch result
pH	N/A	N/A
Conductivity($\mu\text{s}/\text{cm}$)	0 - 10	2.423
Osmolality(mOsm/kg)	N/A	N/A

Preservatives used: No preservatives used

Sterility: Water is UV treated and processed into gamma irradiated bags. Although exposure to the environment is limited, we do not guarantee the sterility of the water.

Storage: Store at Room Temperature

Country of Origin: United Kingdom

Health and Safety:

Precautionary Notes

There are no H&S implications from this product;

Please download the safety data sheet from www.sourcebioscience.com

Signed for Quality Control: 

Date: 15 OCT 18.

Signed for Certificate Check: 

Date: 15 OCT 18