

## CERTIFICATE OF ANALYSIS

# Certificate of Analysis

**Product name:** Heart Valve Cryopreservation Medium

**Product code(s):** 04-311

**Batch Number(s):** 71181011

**Date Manufactured:** 29 Oct 18

**Expiry Date:** 30 Oct 19

**Appearance:** Clear Red Solution

### Quality Control:

	Specification	This batch result
pH	7.0-7.5	7.30
Conductivity( $\mu\text{s}/\text{cm}$ )	N/A	N/A
Osmolality(mOsm/kg)	N/A	N/A

NHS BT Spec  
SPN323/3, Dated 24/09/18.

**Preservatives used:** N/A

**Sterility:** Product is filtered through a 0.2 $\mu\text{m}$  filter and micro tested for sterility.

**Storage:** 0 - 30°C

**Country of Origin:** United Kingdom

### Health and Safety:

#### Precautionary Notes

Please download the Safety Data Sheet at <http://sourcebioscience.com>

**Signed for Quality Control:**



**Date:**

06 Nov 18

**Signed for Certificate Check:**

**Date:**

06 Nov 18