

## CERTIFICATE OF ANALYSIS

# Certificate of Analysis

**Product name:** PBSS (PHOSPHATE BUFFERED SEROLOGY SALINE)

**Product code(s):** 04-004

**Batch Number(s):** 01181159

**Date Manufactured:** 12 Nov 18

**Expiry Date:** 11 Nov 20

**Appearance:** Clear Liquid

**Quality Control:**

	Specification	This batch result
pH	6.95-7.05	7.03
Conductivity( $\mu\text{s}/\text{cm}$ )	N/A	N/A
Osmolality(mOsm/kg)	285-295	289

Meets the requirements set out in 'Guidelines for the Blood Transfusion Services in the United Kingdom' 7th Edition; as set out in chapter 12, 12.2.4, Miscellaneous reagents.

**Preservatives used:** Sodium azide (0.01% [w/v])

**Sterility:** Product is filled into gamma irradiated bags

**Storage:** Room temperature

**Country of Origin:** United Kingdom

**Health and Safety:**

**Precautionary Notes**

Please download the Safety Data Sheet at <http://sourcebioscience.com>

**Signed for Quality Control:**



**Date:** 15 Nov 18

**Signed for Certificate Check:**



**Date:** 15 NOV 18