

## CERTIFICATE OF ANALYSIS

# Certificate of Analysis

**Product name:** PBSS (PHOSPHATE BUFFERED SEROLOGY SALINE )  
**Product code(s):** 04-004  
**Batch Number(s):** 01181182  
**Date Manufactured:** 29 Nov 18  
**Expiry Date:** 28 Nov 20  
**Appearance:** Clear Liquid

### Quality Control:

	Specification	This batch result
pH	6.95-7.05	7.04
Conductivity( $\mu$ s/cm)	N/A	N/A
Osmolality(mOsm/kg)	285-295	292

Meets the requirements set out in 'Guidelines for the Blood Transfusion Services in the United Kingdom' 7th Edition; as set out in chapter 12, 12.2.4, Miscellaneous reagents.

**Preservatives used:** Sodium azide (0.01% [w/v])

**Sterility:** Product is filled into gamma irradiated bags

**Storage:** Room temperature

**Country of Origin:** United Kingdom

### Health and Safety:

#### Precautionary Notes

Please download the Safety Data Sheet at <http://sourcebioscience.com>

Signed for Quality Control:



Date: 03 Dec 18

Signed for Certificate Check:



Date: 03 Dec 18