

## CERTIFICATE OF ANALYSIS

# Certificate of Analysis

**Product name:** PBSS ( PHOSPHATE BUFFERED SEROLOGY SALINE )  
**Product code(s):** 04-004  
**Batch Number(s):** 01190337  
**Date Manufactured:** 07 Mar 19  
**Expiry Date:** 06 Mar 21  
**Appearance:** Clear Liquid  
**Quality Control:**

	Specification	This batch result
pH	6.95-7.05	6.98
Conductivity( $\mu\text{s}/\text{cm}$ )	N/A	N/A
Osmolality(mOsm/kg)	285-295	298

Meets the requirements set out in 'Guidelines for the Blood Transfusion Services in the United Kingdom' 7th Edition; as set out in chapter 12, 12.2.4, Miscellaneous reagents.

**Preservatives used:** Sodium azide (0.01% [w/v])

**Sterility:** Product is filled into gamma irradiated bags

**Storage:** Room temperature

**Country of Origin:** United Kingdom

**Health and Safety:**

### Precautionary Notes

Please download the Safety Data Sheet at <http://sourcebioscience.com>

**Signed for Quality Control:**



**Date:** 12 MAR 19

**Signed for Certificate Check:**



**Date:** 12 MAR 19