

CERTIFICATE OF ANALYSIS

Certificate of Analysis

Product name: PBSS (PHOSPHATE BUFFERED SEROLOGY SALINE)
Product code(s): 04--004/10
Batch Number(s): 01190460
Date Manufactured: 10 Apr 19
Expiry Date: 09 Apr 21
Appearance: Clear Liquid
Quality Control:

	Specification	This batch result
pH	6.95-7.05	6.98
Conductivity(μ s/cm)	N/A	N/A
Osmolality(mOsm/kg)	285-295	293

Meets the requirements set out in 'Guidelines for the Blood Transfusion Services in the United Kingdom' 7th Edition; as set out in chapter 12, 12.2.4, Miscellaneous reagents.

Preservatives used: Sodium azide (0.01% [w/v])

Sterility: Product is filled into Politainer irradiated bags

Storage: Room temperature


Country of Origin: United Kingdom

Health and Safety:

Precautionary Notes

Please download the Safety Data Sheet at <http://sourcebioscience.com>

Signed for Quality Control:



Date: 12 Apr 19

Signed for Certificate Check:



Date: 12 Apr 19