

## CERTIFICATE OF ANALYSIS

# Certificate of Analysis

**Product name:** Phosphate Buffered Serology Saline with EDTA  
**Product code(s):** 04-031  
**Batch Number(s):** 02190403  
**Date Manufactured:** 09 Apr 19  
**Expiry Date:** 16 Jan 20  
**Appearance:** Clear Liquid

**Quality Control:**

	Specification	This batch result
pH	6.95-7.05	6.98
Conductivity( $\mu\text{s}/\text{cm}$ )	N/A	N/A
Osmolality(mOsm/kg)	285-295	287

**Preservatives used:** Sodium azide (0.5g/L)

**Sterility:** Product is filtered through a 0.2 $\mu\text{m}$  filter and tested for micro sterility.

**Storage:** Room temperature

**Country of Origin:** United Kingdom

**Health and Safety:**

**Precautionary Notes**

Please download the Safety Data Sheet at <http://sourcebioscience.com>

**Signed for Quality Control:**



**Date:** 15 Apr 19

**Signed for Certificate Check:**



**Date:** 15 Apr 19