

CERTIFICATE OF ANALYSIS

Certificate of Analysis

Product name: Low Ionic Strength Saline (LISS)

Product code(s): 04-016

Batch Number(s): 03190432

Date Manufactured: 01 Apr 19

Expiry Date: 31 Mar 20

Appearance: Clear Liquid

Quality Control:

	Specification	This batch result
pH	6.50-6.90	6.84
Conductivity($\mu\text{s}/\text{cm}$)	3500-3800	3620
Osmolality(mOsm/kg)	285-295	287

Meets the requirements set out in 'Guidelines for the Blood Transfusion Services in the United Kingdom' 7th Edition; as set out in chapter 12, 12.2.4, Miscellaneous reagents.

Preservatives used: Sodium azide (0.01% [w/v])

Sterility: N/A

Storage: Room temperature

Country of Origin: United Kingdom

Health and Safety:

Precautionary Notes

Please download the Safety Data Sheet at <http://sourcebioscience.com>

Signed for Quality Control: 

Date: 08 APR 19

Signed for Certificate Check: 

Date: 08 APR 19