

CERTIFICATE OF ANALYSIS

Certificate of Analysis

Product name: REAGENT GRADE WATER
Product code(s): 04-035
Batch Number(s): 18190601
Date Manufactured: 26 Jun 19
Expiry Date: 25 Jun 20
Appearance: Clear Liquid
Quality Control:

	Specification	This batch result
pH	N/A	N/A
Conductivity($\mu\text{s}/\text{cm}$)	0 - 10	4.85
Osmolality(mOsm/kg)	N/A	N/A

Preservatives used: No preservatives used in this product

Sterility: water is UV treated and processed into gamma irradiated bags, although exposure to the environment is limited, we do not guarantee the sterility of the water.

Storage: Room temperature

Country of Origin: United Kingdom

Health and Safety:

Precautionary Notes

there are no H&S implications from this product:

please download the safety data sheet at <http://sourcebioscience.com>

Signed for Quality Control: 

Date: 27 Jun 19

Signed for Certificate Check: 

Date: 27 Jun 19