

CERTIFICATE OF ANALYSIS

Certificate of Analysis

Product name: Heart Valve Cryopreservation Medium
Product code(s): 04 - 321
Batch Number(s): 19200101
Date Manufactured: 29 Jan 2020
Expiry Date: 28 Jan 2021
Appearance: Clear Liquid

Quality Control:

	Specification	This batch result
pH	7.20 - 7.40	7.39
Conductivity($\mu\text{s}/\text{cm}$)	9000 - 11000	9618
Osmolality(mOsm/kg)	1950 - 2150	2010

Preservatives used: N/A

Sterility: Product is filtered through a 0.2 μm filter into gamma irradiated bottles

Storage: -20°C

Country of Origin: United Kingdom

Health and Safety:

Precautionary Notes

Please download the Safety Data Sheet at <http://sourcebioscience.com>

Signed for Quality Control: 

Date: 04 Feb 2020

Signed for Certificate Check: 

Date: 04 FEB 2020