

CERTIFICATE OF ANALYSIS

Certificate of Analysis

Product name: Fixol (Fixative Reagent for Fetal Cell Detection Kit)

Product code(s): 05 - 051

Batch Number(s): 51191101

Date Manufactured: 07 Nov 19

Expiry Date: 04 Jul 21

Appearance: Clear Liquid

Quality Control:

	Specification	This batch result
pH	8.5 - 9.5	8.66
Conductivity($\mu\text{s}/\text{cm}$)	1750 - 2150	1810
Osmolality(mOsm/kg)	N/A	N/A

The performance of the above components has been assessed by Kleihauer testing of negative and positive control samples. All tests met the performance specifications for this product and passed the assessment. [Demonstration of fetal hemoglobin in erythrocytes of a blood smear]. KLEIHAUER E, BRAUN H, BETKE K Klin Wochenschr 1957 (35):637-638

Preservatives used: Sodium azide (0.024% [w/v])

Sterility: N/A

Storage: Room temperature

Country of Origin: United Kingdom

Health and Safety:

Precautionary Notes

Please download the Safety Data Sheet at <http://sourcebioscience.com>

Signed for Quality Control:



Date: 12 Feb 2020

Signed for Certificate Check:



Date: 13 Feb 2020