

## CERTIFICATE OF ANALYSIS

# Certificate of Analysis

**Product name:** Fixol (Fixative Reagent for Fetal Cell Detection Kit)

**Product code(s):** 05 - 051

**Batch Number(s):** 51191102

**Date Manufactured:** 13 Nov 19

**Expiry Date:** 01 Sep 21

**Appearance:** Clear Liquid

**Quality Control:**

	Specification	This batch result
pH	8.5 - 9.5	8.95
Conductivity( $\mu\text{s}/\text{cm}$ )	1750 - 2150	1851
Osmolality(mOsm/kg)	N/A	N/A

The performance of the above components has been assessed by Kleihauer testing of negative and positive control samples. All tests met the performance specifications for this product and passed the assessment. [Demonstration of fetal hemoglobin in erythrocytes of a blood smear]. KLEIHAUER E, BRAUN H, BETKE K Klin Wochenschr 1957 (35):637-638

**Preservatives used:** Sodium azide (0.024% [w/v])

**Sterility:** N/A

**Storage:** Room temperature

**Country of Origin:** United Kingdom

**Health and Safety:**

**Precautionary Notes**

Please download the Safety Data Sheet at <http://sourcebioscience.com>

**Signed for Quality Control:** *S.J. Taylor*

**Date:** 12 Feb 2020

**Signed for Certificate Check:** *Waller*

**Date:** 12 FEB 2020