

CERTIFICATE OF ANALYSIS

Certificate of Analysis

Product name: REAGENT GRADE WATER

Product code(s): 04-035

Batch Number(s): 18200201

Date Manufactured: 25 FEB 2020

Expiry Date: 24 FEB 2022

Appearance: Clear Liquid

Quality Control:

	Specification	This batch result
pH	N/a	N/A
Conductivity($\mu\text{s}/\text{cm}$)	Less than 10	3.229
Osmolality(mOsm/kg)	N/a	N/A

Preservatives used: no preservatives used

Sterility: water is UV treated and processed into gamma irradiated bags, although exposure to the environment is limited, we do not guarantee the sterility of the water.

Storage: Room temperature

Country of Origin: United Kingdom

Health and Safety:

Precautionary Notes

there are no H&S implications from this product :please download the safety data sheet at <http://sourcebioscience.com>

Signed for Quality Control:



Date:

02 Mar 2020

Signed for Certificate Check:



Date:

02 MAR 2020