

## CERTIFICATE OF ANALYSIS

# Certificate of Analysis

**Product name:** PBSS (Phosphate Buffered Serology Saline)  
**Product code(s):** 04 - 004/10  
**Batch Number(s):** 01200307  
**Date Manufactured:** 19 Mar 2020  
**Expiry Date:** 18 Mar 2022  
**Appearance:** Clear Liquid  
**Quality Control:**

	Specification	This batch result
pH	6.95-7.05	6.97
Conductivity( $\mu\text{s}/\text{cm}$ )	N/A	N/a
Osmolality(mOsm/kg)	285-295	293

Meets the requirements set out in 'Guidelines for the Blood Transfusion Services in the United Kingdom' 7th Edition; as set out in chapter 12, 12.2.4, Miscellaneous reagents.

**Preservatives used:** Sodium azide (0.01% [w/v])

**Sterility:** Product is filled in Gamma Irradiated bags

**Storage:** Room temperature

**Country of Origin:** United Kingdom

### Health and Safety:

#### Precautionary Notes

Please download the Safety Data Sheet at <http://sourcebioscience.com>

**Signed for Quality Control:**



**Date:** 30 MAR 2020

**Signed for Certificate Check:**



**Date:** 30 MAR 2020