

CERTIFICATE OF ANALYSIS

Certificate of Analysis

Product name: Low Ionic Strength Saline (LISS)
Product code(s): 04-015, 04-016, 04-017
Batch Number(s): 03200401
Date Manufactured: 07 APR 2020
Expiry Date: 21 MAR 2021
Appearance: Clear Liquid

Quality Control:

	Specification	This batch result
pH	6.50 - 6.90	6.77
Conductivity($\mu\text{s}/\text{cm}$)	3500 - 3800	3547
Osmolality(mOsm/kg)	285 - 295	295

Meets the requirements set out in 'Guidelines for the Blood Transfusion Services in the United Kingdom' 7th Edition; as set out in chapter 12, 12.2.4, Miscellaneous reagents.

Preservatives used: Sodium azide (0.1% [w/v])

Sterility: Product is filled into gamma irradiated bags and bottles

Storage: Room temperature

Country of Origin: United Kingdom

Health and Safety:

Precautionary Notes

Please download the Safety Data Sheet at <http://sourcebioscience.com>

Signed for Quality Control:



Date:

20 APR 2020

Signed for Certificate Check:



Date:

20 Apr 2020