

CERTIFICATE OF ANALYSIS

Certificate of Analysis

Product name: PBSS (Phosphate Buffered Serology Saline)
Product code(s): 04 - 004
Batch Number(s): 01200401
Date Manufactured: 07 APR 2020
Expiry Date: 06 APR 2022
Appearance: Clear Liquid

Quality Control:

| | Specification | This batch result |
|---|---------------|-------------------|
| pH | 6.95-7.05 | 6.96 |
| Conductivity($\mu\text{s}/\text{cm}$) | N/A | N/a |
| Osmolality(mOsm/kg) | 285-295 | 293 |

Meets the requirements set out in 'Guidelines for the Blood Transfusion Services in the United Kingdom' 7th Edition; as set out in chapter 12, 12.2.4, Miscellaneous reagents.

Preservatives used: Sodium azide (0.01% [w/v])

Sterility: Product is filled in Gamma Irradiated bags

Storage: Room temperature

Country of Origin: United Kingdom

Health and Safety:

Precautionary Notes

Please download the Safety Data Sheet at <http://sourcebioscience.com>

Signed for Quality Control:



Date: 20 Apr 2020

Signed for Certificate Check:



Date: 20 APR 2020