

CERTIFICATE OF ANALYSIS

Certificate of Analysis

Product name: Tris Buffered Isotonic Saline (TBIS)

Product code(s): 04 - 142

Batch Number(s): 37200601

Date Manufactured: 17 Jun 2020

Expiry Date: 02 Jun 2021

Appearance: Clear Fluid

Quality Control:

	Specification	This batch result
pH	7.3 - 7.6	7.48
Conductivity($\mu\text{s}/\text{cm}$)	14000 - 16500	14207
Osmolality(mOsm/kg)	285 - 300	296

Manufactured to SPN 437/1 Effective 11May12

Preservatives used: N/A

Sterility: Product is filtered through a 0.2 μm filter into Gamma Irradiated Bottles

Storage: 0 - 30 °C

Country of Origin: United Kingdom

Health and Safety:

Precautionary Notes

Please download the Safety Data Sheet at <http://sourcebioscience.com>

Signed for Quality Control:



Date: 24 Jun 2020

Signed for Certificate Check:



Date: 24 JUN 2020