

CERTIFICATE OF ANALYSIS

Certificate of Analysis

Product name: Detergent Buffer Base

Product code(s): 04 - 334

Batch Number(s): 77200701

Date Manufactured: 08 Jul 2020

Expiry Date: 07 Jul 2021

Appearance: Clear Solution

Quality Control:

	Specification	This batch result
pH	7.8 - 8.2	8.1
Conductivity($\mu\text{s}/\text{cm}$)	N/A	N/A
Osmolality(mOsm/kg)	N/A	N/A

Manufactured to SPN 321/2 Effective 05Apr17

Preservatives used: N/A

Sterility: Product is filtered through a 0.2 μm filter into Gamma Irradiated Bottles

Storage: 2 - 25°C

Country of Origin: United Kingdom

Health and Safety:

Precautionary Notes

Please download the Safety Data Sheet at <http://sourcebioscience.com>

Signed for Quality Control: *S. J. Taylor*

Date: *13 Jul 2020*

Signed for Certificate Check: *Prnya Gordon*

Date: *13 JUL 2020*