

CERTIFICATE OF ANALYSIS

Certificate of Analysis

Product name: Decellularised Dermis Cryoprotectant Medium (DCD Cryo)
Product code(s): 04 - 351
Batch Number(s): 83200701
Date Manufactured: 06 Jul 2020
Expiry Date: 17 Jun 2021
Appearance: Clear Solution
Quality Control:

	Specification	This batch result
pH	7.10 - 7.50	7.48
Conductivity($\mu\text{s}/\text{cm}$)	N/A	N/A
Osmolality(mOsm/kg)	N/A	N/A

Manufactured to 430/3 Effective 19OCT17

Preservatives used: N/A

Sterility: Product is filtered through a 0.2 μm filter into Gamma Irradiated Bottles

Storage: 2 - 25°C

Country of Origin: United Kingdom

Health and Safety:

Precautionary Notes

Please download the Safety Data Sheet at <http://sourcebioscience.com>

Signed for Quality Control: *S.J. Taylor*

Date: 14 JUL 2020

Signed for Certificate Check: *Priya Gordon*

Date: 14 JUL 20