

## CERTIFICATE OF ANALYSIS

# Certificate of Analysis

**Product name:** REAGENT GRADE WATER  
**Product code(s):** 04-035  
**Batch Number(s):** 18200702  
**Date Manufactured:** 16 JULY 2020  
**Expiry Date:** 15 JULY 2021  
**Appearance:** Clear Liquid  
**Quality Control:**

	Specification	This batch result
pH	N/a	N/A
Conductivity( $\mu\text{s}/\text{cm}$ )	Less than 10	5.817
Osmolality(mOsm/kg)	N/a	N/A

**Preservatives used:** no preservatives used

**Sterility:** water is UV treated and processed into gamma irradiated bags, although exposure to the environment is limited, we do not guarantee the sterility of the water.


**Storage:** Room temperature

**Country of Origin:** United Kingdom

### Health and Safety:

#### Precautionary Notes

there are no H&S implications from this product :please download the safety data sheet at <http://sourcebioscience.com>

**Signed for Quality Control:** 

**Date:** 20 July 2020

**Signed for Certificate Check:** Priya Gordon

**Date:** 20 JUL 20