

CERTIFICATE OF ANALYSIS

Certificate of Analysis

Product name: Phosphate Buffered Serology Saline

Product code(s): 04-004

Batch Number(s): 01201001

Date Manufactured: 09 Oct 20

Expiry Date: 08 Oct 22

Appearance: Clear Liquid

Quality Control:

	Specification	This batch result
pH	6.95 – 7.05	7.01
Osmolality(mOsm/kg)	285 - 295	287

Meets the requirements set out in 'Guidelines for the Blood Transfusion Services in the United Kingdom' 7th Edition; as set out in chapter 12, 12.2.4, Miscellaneous reagents.

Preservatives used: Sodium azide (0.01% [w/v])

Sterility: Product is filled in Gamma Irradiated bags

Storage: Room temperature

Country of Origin: United Kingdom

Health and Safety:

Precautionary Notes

2Please download the Safety Data Sheet at <http://sourcebioscience.com>

Signed for Quality Control:

Date: 13 Oct 2020

Signed for Certificate Check:

Date: 13 Oct 2020

Author: Thomas Waller
Technical Reviewer: Sam Taylor
QA Reviewer: Jahira Jangaria

SER-COA-21 v1.0
Authorised Date: 23 Jun 20
Next Review Date: 23 Jun 22