

## CERTIFICATE OF ANALYSIS

# Certificate of Analysis

**Product name:** Phosphate Buffered Serology Saline

**Product code(s):** 04-004

**Batch Number(s):** 01201007

**Date Manufactured:** 28 Oct 20

**Expiry Date:** 27 Oct 22

**Appearance:** Clear Liquid

### Quality Control:

	Specification	This batch result
pH	6.95 – 7.05	7.03
Osmolality(mOsm/kg)	285 - 295	291

Meets the requirements set out in 'Guidelines for the Blood Transfusion Services in the United Kingdom' 7th Edition; as set out in chapter 12, 12.2.4, Miscellaneous reagents.

**Preservatives used:** Sodium azide (0.01% [w/v])

**Sterility:** Product is filled in Gamma Irradiated bags

**Storage:** Room temperature

**Country of Origin:** United Kingdom

### Health and Safety:

#### Precautionary Notes

Please download the Safety Data Sheet at <http://sourcebioscience.com>

Signed for Quality Control:



Date: 03 NOV 2020

Signed for Certificate Check:



Date: 03 NOV 2020