

CERTIFICATE OF ANALYSIS

# Certificate of Analysis

**Product name:** Phosphate Buffered Serology Saline

**Product code(s):** 04-004

**Batch Number(s):** 01201106

**Date Manufactured:** 24 Nov 2020

**Expiry Date:** 23 Nov 2022

**Appearance:** Clear Liquid

**Quality Control:**

	Specification	This batch result
pH	6.95 – 7.05	7.02
Osmolality(mOsm/kg)	285 - 295	294

Meets the requirements set out in 'Guidelines for the Blood Transfusion Services in the United Kingdom' 7th Edition; as set out in chapter 12, 12.2.4, Miscellaneous reagents.

**Preservatives used:** Sodium azide (0.01% [w/v])

**Sterility:** Product is filled in Gamma Irradiated bags

**Storage:** Room temperature

**Country of Origin:** United Kingdom


**Health and Safety:**

**Precautionary Notes**

Please download the Safety Data Sheet at <http://sourcebioscience.com>

**Signed for Quality Control:** 

**Date:** 26 Nov 2020

**Signed for Certificate Check:** 

**Date:** 26 Nov 2020