

CERTIFICATE OF ANALYSIS

Certificate of Analysis

Product name: 0.5M Hydrochloric acid
Product code(s): 04-356
Batch Number(s): 84201101
Date Manufactured: 26.Nov 2020
Expiry Date: 13 Feb 2022
Appearance: Clear Liquid
Quality Control:

	Specification	This batch result
Osmolality(mOsm/kg)	920-1020	986

Manufactured to SPN 435/1 – Effective date 28 Mar 2012

Preservatives used: N/a

Sterility: Product is filtered through a 0.2µm filter into gamma irradiated bottles.

Storage: Ambient

Country of Origin: United Kingdom

Health and Safety:

Precautionary Notes

Please download the Safety Data Sheet at <http://sourcebioscience.com>

Signed for Quality Control:



Date: 02 Dec 2020

Signed for Certificate Check:



Date: 02 Dec 2020