

CERTIFICATE OF ANALYSIS

Certificate of Analysis

Product name: Low Ionic Strength Saline (LISS)

Product code(s): 04 - 015 & 04-016,

Batch Number(s): 03210101

Date Manufactured: 06 Jan 2021

Expiry Date: 05 Jan 2022

Appearance: Clear Liquid

Quality Control:

	Specification	This batch result
pH	6.50 - 6.90	6.85
Conductivity($\mu\text{s}/\text{cm}$)	3500 - 3800	3635
Osmolality(mOsm/kg)	285 - 295	290

Meets the requirements set out in 'Guidelines for the Blood Transfusion Services in the United Kingdom' 7th Edition; as set out in chapter 12, 12.2.4, Miscellaneous reagents.

Preservatives used: Sodium azide (0.01% [w/v])

Sterility: N/A

Storage: Room temperature

Country of Origin: United Kingdom

Health and Safety:

Precautionary Notes

Please download the Safety Data Sheet at <http://sourcebioscience.com>

Signed for Quality Control:



Date:

11 JAN 2021

Signed for Certificate Check:



Date:

11 Jan 2021