

CERTIFICATE OF ANALYSIS

Certificate of Analysis

Product name: Phosphate Buffered Serology Saline

Product code(s): 04-004/10

Batch Number(s): 01210106

Date Manufactured: 21 Jan 2021

Expiry Date: 20 Jan 2023

Appearance: Clear Liquid

Quality Control:

	Specification	This batch result
pH	6.95 – 7.05	7.04
Osmolality(mOsm/kg)	285 - 295	289

Meets the requirements set out in 'Guidelines for the Blood Transfusion Services in the United Kingdom' 7th Edition; as set out in chapter 12, 12.2.4, Miscellaneous reagents.

Preservatives used: Sodium azide (0.01% [w/v])

Sterility: Product is filled in Gamma Irradiated bags

Storage: Room temperature

Country of Origin: United Kingdom

Health and Safety:

Precautionary Notes

Please download the Safety Data Sheet at <http://sourcebioscience.com>

Signed for Quality Control: 

Date: 22 Jan 21

Signed for Certificate Check: 

Date: 22 JAN 21