

## CERTIFICATE OF ANALYSIS

# Certificate of Analysis

**Product name:** Phosphate Buffered Serology Saline

**Product code(s):** 04-004

**Batch Number(s):** 01210202

**Date Manufactured:** 08 Feb 2021

**Expiry Date:** 07 Feb 2023

**Appearance:** Clear Liquid

**Quality Control:**

|                     | Specification | This batch result |
|---------------------|---------------|-------------------|
| pH                  | 6.95 – 7.05   | 6.97              |
| Osmolality(mOsm/kg) | 285 - 295     | 291               |

Meets the requirements set out in 'Guidelines for the Blood Transfusion Services in the United Kingdom' 7th Edition; as set out in chapter 12, 12.2.4, Miscellaneous reagents.

**Preservatives used:** Sodium azide (0.01% [w/v])

**Sterility:** Product is filled in Gamma Irradiated bags

**Storage:** Room temperature

**Country of Origin:** United Kingdom

**Health and Safety:**

**Precautionary Notes**

Please download the Safety Data Sheet at <http://sourcebioscience.com>

**Signed for Quality Control:** 

**Date:** 08 Feb 21

**Signed for Certificate Check:** 

**Date:** 08 FEB 21