

CERTIFICATE OF ANALYSIS

Certificate of Analysis

Product name: Cryoprotectant Reagent

Product code(s): 04 - 213

Batch Number(s): 61210201

Date Manufactured: 17 Feb 2021

Expiry Date: 13 Feb 2022

Appearance: Clear Liquid

Quality Control:

	Specification	This batch result
pH	5.80 - 6.70	6.40
Conductivity($\mu\text{s}/\text{cm}$)	4200 - 4850	4317
Osmolality(mOsm/kg)	880 - 970 (when diluted 1:5 with water)	892

Preservatives used: N/A

Sterility: Product is filtered through a 0.2 μm into gamma irradiated bags

Storage: Ambient

Country of Origin: United Kingdom

Health and Safety:

Precautionary Notes

Please download the Safety Data Sheet at <http://sourcebioscience.com>

Signed for Quality Control: 

Date: 24 FEB 21

Signed for Certificate Check: 

Date: 24 FEB 21