

CERTIFICATE OF ANALYSIS

Certificate of Analysis

Product name: Fixol (Fixative Reagent for Fetal Cell Detection Kits)

Product code(s): 05 - 051

Batch Number(s): 51210201

Date Manufactured: 25 Feb 2021

Expiry Date: 20 Feb 2023

Appearance: Clear Liquid

Quality Control:

	Specification	This batch result
pH	8.5 - 9.5	8.88
Conductivity($\mu\text{s}/\text{cm}$)	1750 - 2150	1874
Osmolality(mOsm/kg)	N/A	N/A

The performance of the above components has been assessed by Kleihauer testing of negative and positive control samples. All tests met the performance specifications for this product and passed the assessment. [Demonstration of fetal hemoglobin in erythrocytes of a blood smear]. KLEIHAUER E, BRAUN H, BETKE K Klin Wochenschr 1957 (35):637-638

Preservatives used: Sodium Azide (0.024% w/v)

Sterility: N/A

Storage: Room Temperature

Country of Origin: United Kingdom

Health and Safety:

Precautionary Notes

Please download the Safety Data Sheet at <http://sourcebioscience.com>

Signed for Quality Control:



Date: 03 MAR 21

Signed for Certificate Check:



Date: 03 MAR 21