

CERTIFICATE OF ANALYSIS

Certificate of Analysis

Product name: Reagent grade water

Product code(s): 04 - 035

Batch Number(s): 18210302

Date Manufactured: 26 March 2021

Expiry Date: 25 March 2023

Appearance: Clear Liquid

Quality Control:

| | Specification | This batch result |
|---|---------------|-------------------|
| Conductivity($\mu\text{s}/\text{cm}$) | < 10 | 2.33 |
| Osmolality(mOsm/kg) | < 10 | -5 |

Preservatives used: N/a

Sterility: Product is dispensed into gamma irradiated bags

Storage: Ambient

Country of Origin: United Kingdom

Health and Safety:

Precautionary Notes

Please download the Safety Data Sheet at <http://sourcebioscience.com>

Signed for Quality Control: 

Date: 29 MAR 21

Signed for Certificate Check: 

Date: 29 MAR 21